



KEY FEATURES	BENEFITS	
Low Stray Light:	New convergent light source improves accuracy and precision	
Fast:	Acquisition, reconstruction and display of optical CT images in approximately 5 minutes	
Efficient:	Simple on-screen controls and straightforward workflow	
Flexible:	Different wavelengths available for different dosimeters, interchangeable aquariums available for different index of refraction matching liquids, open accessible data for research	
Compatible:	Works with ClearView [™] and 3 rd party radiochromic 3D dosimeters	
Clean:	Top load design with dosimeter clamp on rotary stage reduces operator contact with dosimeters and aquarium liquids	
Two Sizes:	For large or small 3D dosimeters	

Existing laser scanners are slow, complicated and not clinically practical. Sparsely sampled 2D detector arrays do not provide enough points for confidence in 3D.

Vista[™] 16 is a fast, reliable, geometrically accurate and dosimetrically accurate cone beam optical CT scanner.

Designed for *True 3D* dosimetry, Vista[™]16 is used for on-site or remote scanning of ClearView[™] 3D dosimeters. The optical CT scanner includes software for acquisition, reconstruction, and display of 3D optical CT images. Images from Vista[™] 16 are analyzed in a variety of 3rd party software applications.

Vista[™] 16 makes use of a new convergent light source design which has been shown to reduce stray light and improve the accuracy and precision of optical CT measurements⁽¹⁾.

The Vista[™] 16 optical CT scanner is designed to work with radiochromic dosimeters, such as Fricke which have an absorption peak at 590 nm or 633 nm. Different wavelengths are available on request for different dosimeters and applications.



Above: Close-up of the liquid filled Vista[™]16 aquarium capable of scanning objects up to 15 cm diameter by 12 cm long.

1: Dekker KH, Battista JJ, Jordan KJ. Stray light in cone beam optical computed tomography: II. Reduction using a convergent light source.

Phys Med Biol. 2016 Apr 7;61(7):2910-25. doi: 10.1088/0031-9155/61/7/2910. PubMed PMID: 26988107.



Vista[™] 16 is capable of scanning radiochromic dosimeters up to 15 cm diameter by 12 cm long. Objects are suspended in a liquid filled aquarium to provide index of refraction matching. Data is fully accessible for research applications.

Performance	Typical Scanning Laser Systems	Vista Optical CT Scanner
Scanning Time	6 minutes per slice	Less than 5 minutes for 360 projections, up to 400 slices
Reconstruction Time	6 minutes per slice	Less than 5 minutes for 100 slices
Spatial Resolution	0.5 mm, non isotropic	2.0, 1.0, 0.5, 0.25 mm, isotropic
Overall Time to Results	4 hours – 1 day or more	Less than 1 hour

Above: *Results based on 1. M. Oldham, J. H. Siewerdsen, S. Kumar, J. Wong and D. A. Jaffray; Optical-CT gel-dosimetry I: Basic investigations, Med. Phys. 30, 623-634, 2003



Above: Suface and color rendering of a multi-beam dose distribution.

ORDERING INFORMATION

100-2016 Vista[™] 16 Cone Beam Optical CT Scanner System Includes:

- ► Scanner
- Software
- User's Guide
- Aquarium for index of refraction matching liquid
- Wavelength 530 nm (other wavelengths available on request)
- ► 15 cm Rotary Stage
- 5 Polyethylene terephthalate jars (15 cm diameter) Other Rotary Stage and Jar sizes available on request.

SPECIFICATIONS

- Aquarium:
 Glass/Aluminum
- Region of interest dimensions:
 Up to 15 cm diameter x 12 cm long
- Enclosure (Height includes cabinet containing motor):
 35 cm x 165 cm x 48 cm (WxLxH)
- Electrical 24V DC, 2.1A ; 120 240 V AC medical grade switching power supply provided

ACCESSORIES

500-1067	Annual Service & Software Support
500-1016	Aquarium (For Index of Refraction Matching)
500-1045	Additional Light Source (Wavelength to be
	confirmed prior to ordering)
500-1072	10 cm Rotary Stage
500-1073	Dovetail Rotary Stage and Jar Clamp for
	Smaller Jars
500-	-1074 Jar Clamp for Smaller jars

MINIMUM TECHNICAL REQUIREMENTS

- Operating System: Windows 7, 10
- ▶ Processor: i7, 2 GHz or better
- ▶ **RAM**: 4GB or more
- ► Graphics Card: 512 MB or better with OpenGL 2.1
- Peripheral Connectivity: USB 3.0

© 2020 Modus Medical Devices Inc. All Rights Reserved. Specifications subject to change without notice. Modus QA is not responsible for errors or omissions.

GET SOCIAL

í in

Modus QA makes no warranties or commitments concerning the availability of future products or versions that may be planned or under development. PDS #100-2016, REV#01.20

Modus QA

1570 North Routledge Park, London, Ontario Canada N6H 5L6

[e] info@modusQA.com[w] www.modusQA.com

Toll Free: +1 (866) 862-9682 (North America) Phone: +1 (519) 438-2409 Fax: +1 (519) 643-0127

